



Water Testing HB (2016) Limited

Client Number 9004

PO Box 2552, Stortford Lodge, Hastings, 4153
1105 Plunket Street, Saint Leonards, Hastings, 4120

Telephone 06 870-6449

www.watertestinghb.nz

Authorised Representative

Ms Cynthia Bowers
General Manager

Programme

Chemical Testing Laboratory

Accreditation Number 1360

Initial Accreditation Date 8 July 2020

Conformance Standard

ISO/IEC 17025:2017

General requirements for the competence of testing and calibration laboratories

Laboratory Services Summary

Analysis

2.41 Waters

Sampling

2.41 Waters

Key Technical Personnel

Analysis

Ms Aania Anderson 2.41
Mr Caleb Hart 2.41

Sampling

Ms Aania Anderson 2.41
Mr Caleb Hart 2.41

Operations Manager Authorisation:		Issue 3	Date:07/08/20	Page 1 of 2
--------------------------------------	--	---------	---------------	-------------



Water Testing HB (2016) Limited
 Chemical Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation Number 1360

Analysis

2.41 Waters

- (a) Potable waters
- (b) Non-potable waters

Free available chlorine	APHA 4500-Cl G
pH	APHA 4500-H B
Total chlorine	APHA 4500-Cl G
Turbidity	APHA 2130 B

- (a) Potable waters
- (b) Non-potable waters (e.g. receiving waters, ground waters)
- (d) Effluents and trade wastes
- (f) Swimming and spa pools
- (g) Marine waters

Temperature	APHA 2550 B
-------------	-------------


Sampling

2.41 Waters

- (a) Potable waters
- (b) Non-potable waters
- (d) Effluents and trade wastes
- (f) Swimming pools and spas
- (g) Marine waters

In accordance with in-house methods based on the following:

- A National Protocol for State of the Environment Groundwater Sampling in New Zealand
- US Environmental Protection Agency August 2015, Science and Ecosystem Support Division, "Operating Procedure: Field Equipment Cleaning and Decontamination"
- A. Pita. 2015. Sampling Instructions for USEPA 1623 (modified). Hopkirk Institute, Massey University, Palmerston North
- Standard Methods for the Examination of Water and Wastewater 23rd Edition

Operations Manager Authorisation:		Issue 3	Date:07/08/20	Page 2 of 2
--------------------------------------	---	---------	---------------	-------------